



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 8:06 am, Jun 04, 2013

REPORT#6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|--|--|----------------------------------|
| DATAMASTER SN 201206 | NAME OF AGENCY St. Joseph Police Department | DATE OF INSPECTION 06-03-2013 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 501 Faraon Street St. Joseph, MO 64501 | | TIME OF INSPECTION 1734 |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

| | |
|---|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) | DATE AND TIME (from printout) 06/03/13 17:34 |
| <input checked="" type="checkbox"/> COMPUTER | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTERS |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 oc | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR | <input checked="" type="checkbox"/> CALIBRATION |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED | <input checked="" type="checkbox"/> PRINTER |
| <input checked="" type="checkbox"/> INDICATOR LIGHTS | |
| <input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Repco Marketing LOT # 12001 EXRDATE 04-02-2014 | |
| <input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 oc SIMULATOR SN SD2278 EXP. DATE 12-17-2013 | |
| <input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) | |
| Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) | |
| <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE | |
| <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE | |
| <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE | |

| | | |
|----------------|----------------|---------------|
| TEST 1 w- .100 | TEST 2 a- .100 | TEST 3 w .100 |
|----------------|----------------|---------------|

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | |
|------------|-----------|-------------|-------------|-------------|-----------|
| REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | OVER.19 1 |
|------------|-----------|-------------|-------------|-------------|-----------|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

| | |
|--|--------------------------------------|
| SIGNATURE | PRINT FULL NAME Sgt. Chris McBane |
| TYPE 11 PERMIT NUMBER/EXPIRATION DATE 220199 08-13-2014 | TELEPHONE NUMBER (816) 271-4712 |

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12001

EXPIRATION DATE: April 2, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1205 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is April 3, 2012.
The expiration date for this lot number is April 2, 2014
at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
ST. JOSEPH POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201206
06/03/13

TESTING OFFICER:
MCBRANE/C
OFFICER I.D.: 3740
PERMIT NUMBER: 220199
EXPIRATION DATE: 08/13/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

| | | |
|-------------------|----------|-------|
| BLANK TEST | .000 | 17:36 |
| INTERNAL STANDARD | VERIFIED | 17:36 |
| EXTERNAL STANDARD | .100 | 17:37 |
| BLANK TEST | .000 | 17:37 |
| EXTERNAL STANDARD | .100 | 17:38 |
| BLANK TEST | .000 | 17:38 |
| EXTERNAL STANDARD | .100 | 17:39 |
| BLANK TEST | .000 | 17:39 |

N = 3
SIM. = .1
RVB. = .1

OPERATOR SIGNATURE

Sgt. M. M. #3740

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
ST. JOSEPH POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201206
06/03/13

ARREST TIME: 17:00
SUBJECT NAME:
RFI/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/12345
ARRESTING OFFICER:

MCBRANE/C
OFFICER I.D.: 3740
TESTING OFFICER:

MCBRANE/C
OFFICER I.D.: 3740
PERMIT NUMBER: 220199
EXPIRATION DATE: 08/13/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

| | | |
|--------------------|----------|-------|
| BLANK TEST | .000 | 17:41 |
| INTERNAL STANDARD | VERIFIED | 17:41 |
| RADIO INTERFERENCE | | |

OPERATOR SIGNATURE

Sgt. M. M. #3740

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
ST. JOSEPH POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201206

06/03/13
17:34

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~

OPERATOR SIGNATURE

Spencer M. #3740

Card Stock No.
60021

REORDER ALL SUPPLIES FROM NPAS.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



CHRISTOPHER MCBANE

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 08/13/2012

Number 220199

Expires 08/13/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)